

Pre Bid meeting held on 27.04.2016 under the Chairmanship of Director, I.N.U at the chamber of Director, I.N.U regarding e-Tender invited for the Expression of Interest for Providing Medical Equipments under Public Private Partnership

Members Present:

- 1) Dr. R. Keshavamurthy
Director
I.N.U, Bengaluru. Chairman
- 2) Dr. Umesha L.
Medical Superintendent
I.N.U, Bengaluru. Member
- 3) Dr. Asha .S.
Administrative Officer
I.N.U, Bengaluru. Member Secretary

Special Invitee :

- 4) Dr. Ramalingaiah
Associate Professor of Radiology
I.N.U, Bengaluru.

The representatives of the following Agency were present in the meeting.

Sl. No.	Name of the Agency	Name of the Representative
1.	Blue Star Engineering & Electronic Ltd.	Mr. P. Baskar

Member Secretary welcomed the participants and requested to commence the meeting.

The Representative of the prospective bidders queried/ sought clarification verbally/written on various issues which were clarified by the members of the committee present. The details are as follows:

Procurement of C.T. Scan Machine:

Sl. No.	Name of the prospective Bidder/ Representative	Tender Document reference	Queries	Clarifications
1	M/s.Blue Star Engineering & Electronic Ltd. Mr. P. Baskar	Page No. 40 Image Reconstruction Recon time - Should have 20 images /Sec/ more	Reconstruction time of 20 images/Sec is specific to a single company. Most of the 16 slice C.T. Machine available in the market have a Reconstruction time start with min10 images/Sec. This will not effect the quality of real time images and performance of the machine. Hence requested to amend Reconstruction time to 10 images/Sec or more.	Committee after the discussion decided to amend the Reconstruction time to 10 images/Sec or more instead of 20 images/Sec/more.
2		Page No. 41 After Sales Service: a) After sales service centre should be available in India and 95% uptime service. The service should be provided directly by main manufacturer.	Requested to amend to “The service should be provided by the manufacturer or authorized dealer” instead of “The service should be provided directly by main manufacturer” .	After the discussion the committee decided to amend as “The service should be provided directly by manufacturer or authorized dealer” .

At the conclusion, the participants were informed that the responses to the pre bid queries would be hosted on the e-portal and website of the institute at the earliest.

Meeting concluded with vote of thanks.

Sd/-
Director,
Institute of Nephro-Urology,
Bangalore-02.